

DRAFT
DOCUMENT

DD 1
(Inf.)

43rd CIML Meeting

Sydney 2008 (Item 6.3)

INFORMATION

New Document

Guide for the application of ISO IEC Guide 65
to assessment of measuring instrument certification bodies
in legal metrology

Results of the postal ballot and comments received




ORGANISATION INTERNATIONALE
DE MÉTROLOGIE LÉGALE

INTERNATIONAL ORGANIZATION
OF LEGAL METROLOGY

Project Number 48 (Guide for the application of ISO IEC Guide 65 to assessment of measuring instrument certification bodies in legal metrology)

Deadline: 2008-09-03

AUSTRIA voted Yes
 BELARUS voted Yes
 BRAZIL voted Yes
 CANADA voted Yes
 CYPRUS voted Yes
 CZECH REPUBLIC voted Yes
 DENMARK voted Yes
 FINLAND voted Yes
 GERMANY voted Yes
 IRAN voted Yes
 JAPAN voted Yes (Comments)
 NETHERLANDS voted Yes
 P.R. CHINA voted Yes
 POLAND voted Yes
 SERBIA voted Yes
 SLOVAKIA voted Yes
 SOUTH AFRICA voted Yes
 SWEDEN voted Yes
 SWITZERLAND voted Yes
 UNITED KINGDOM voted Yes
 UNITED STATES voted Yes



Countries who did not vote (38)

ALBANIA, ALGERIA, AUSTRALIA, BELGIUM, BULGARIA, CAMEROON, CROATIA, CUBA, EGYPT, ETHIOPIA, FRANCE, GREECE, HUNGARY, INDIA, INDONESIA, IRELAND, ISRAEL, ITALY, KAZAKHSTAN, KENYA, KOREA (R.), MACEDONIA (F.Y.R.), MONACO, MOROCCO, NEW ZEALAND, NORWAY, PAKISTAN, PORTUGAL, ROMANIA, RUSSIAN FEDERATION, SAUDI ARABIA, SLOVENIA, SPAIN, SRI LANKA, TANZANIA, TUNISIA, TURKEY, VIET NAM.

Synthesis of comments received on the draft OIML Document

Guide for the application of ISO IEC Guide 65 to assessment of measuring instrument certification bodies in legal metrology

Draft Document	Comments on and proposals for modifications to the second column of the DD table	Secretariat replies
Introduction		
4 Certification body		
4.2 Organization	<p>G.4.2-2</p> <p>In case of regulatory certification, withdrawing and/or suspending certifications shall be done in accordance/cooperation with the regulatory designating authority.</p> <p>< Comments ></p> <p>Standpoint of the regulatory designating authority is not always equal to that of the certification body.</p>	<p>Accepted. Even if the meaning of the Guidance is what you mentioned in your comment, we suggest the following clarification.</p> <p><i>Proposal:</i> <i>In case of regulatory certification, withdrawing and/or suspending certifications shall be done in cooperation with the regulatory designating authority and in accordance with the rules defined by the regulatory designating authority.</i></p>
4.3 Operations	<p>G.4.3-3</p> <p>"When this document is used in relation to the OIML basic certificate (or MAA certificate)" is to be added at the beginning of this paragraph.</p> <p>< Comments ></p> <p>In order to clearly state that application of ISO/IEC 17025 in ISO/IEC Guide 65 is limited to B3 (or B10).</p>	<p>Not accepted. The use of this OIML Guide is not limited to the implementation of OIML B 3 and OIML B 10. It may also be used by national authorities to evaluate type approval bodies, for instance when the national regulation refers to ISO/IEC Guide 65 criteria for designating national type approval bodies. In such a case, national type approval bodies are responsible for evaluating that their subcontracting testing laboratories conform to ISO/IEC 17025.</p>

Draft Document	Comments on and proposals for modifications to the second column of the DD table	Secretariat replies
<p>4.4</p> <p>Subcontracting</p>	<p>G.4.4-2</p> <p>"When this document is used in relation to the OIML Basic Certificate (or MAA Certificate)" is to be added at the beginning of this paragraph.</p> <p>< Comments ></p> <p>In order to clearly state that application of ISO/IEC 17025 in ISO/IEC Guide 65 is limited to B3 (or B10).</p>	<p>Not accepted. See reply to 4.3.</p>